



中国认可  
国际互认  
检测  
TESTING  
CNAS L10118



**国检检测**  
CHINA COMPONENTS TEST

# Test Report

(2020) WSZ FHL NO.2275

Product Name KN95 3D Mask

Applicant Ho Dun Goldshiele (Shanghai) Ltd

Manufacturer Suzhou letian protective products Co., Ltd

Test Type Entrusted inspection





Jiangsu Guojian Testing Technology Co.,Ltd



# Test Report

[2020] WSZ FHL NO.2275

Page 1 of 2

Product name	KN95 3D Mask	Specification	G035
		Brand	Ho Dun
Applicant/Add /Tel	Ho Dun Goldshiele (Shanghai) Ltd/Room 331,Build 2,No.15,Riyang South Road , China (Shanghai) Pilot Free Trade Zone/22310590		
Manufacturer/ Add/Tel	Suzhou letian protective products Co., Ltd/12 Zhengfeng Road,Nanfeng Town,Zhangjiagang City/—		
Sample grade	KN95	Sample number	GW 2275-2020
Sample quantity	10PCS	Receiving date of sample	18/03/2020
Test type	Entrusted inspection	Article number/Batch number/Style number	—
Test date	18/03/2020~21/03/2020	Test site	Testing room
Sample state	Meet the requirements		
Test standard(s)	GB 2626-2006 《Respiratory protective equipment -Non-powered purifying particle respirator》		
Test items	Filter efficiency、 Respiratory resistance		
Test conclusion	<p>The sample upon examination, the projects meet the requirements of the GB 2626-2006 standard , the specific test results see page 2.</p> 		
Note	<p>The sample information are provided by Applicant. This report is only responsible for samples received.</p>		

Approver: 苏晓群

Reviewer: 万恒

Main surveyor: 杨莹

# Test Result

[2020] WSZ FHL NO.2275

Page 2 of 2

Number	Test item	Unit	Technical requirements	Test Result		Single Item decision
1	Filter efficiency	—	$\geq 95.0\%$	Untreated	99.5%	Qualified
					99.5%	
					99.6%	
					99.4%	
					99.5%	
					99.4%	
					99.5%	
				Pretreatment	99.4%	
					99.3%	
					99.2%	
					99.6%	
					99.4%	
					99.3%	
					99.5%	
2	Respiratory resistance	Pa	$\leq 350$	Untreated	113.7	Qualified
					116.8	
				Pretreatment	111.6	
					113.7	
	Expiratory resistance	Pa	$\leq 250$	Untreated	94.1	Qualified
					96.2	
				Pretreatment	91.3	
					92.7	
Note						

The End